

i-Cushion

Instructions for Use

DESCRIPTION

The package consists of individual sterile disposable i-Cushion for single use only.

INDICATION

- Only to be used with I medical's Lens delivery system.

MODE OF ACTION

- Acts as a cushion between lens and plunger to prevent the surface quality of intraocular lens

CONTRAINDICATION

- Patient with a known allergy or intolerance to device material

PRECAUTIONS

- Contents are sterile unless the package is opened or damaged.
- The package is ETO sterilized and must be opened only under aseptic conditions.
- Do not reuse the cushion
- Always ensure that the i-Cushion is fully fitted on the plunger tip
- Do not autoclave this product
- This product should be stored at room temperature
- Use of this product other then intended purpose is not recommended

DIRECTION FOR USE

- Prior to opening the pouch, verify the label for correct model and expiration date
- Inspect the pouch carefully before opening
- In case tears, cuts, punctures, or other similar sign is observed, chances are that the pouch is opened or damaged
- To remove the i-Cushion, open the pouch and transfer it to a sterile environment. Carefully fit the i-Cushion on the tip of the plunger

HOW SUPPLIED

i-Cushion is supplied sterile in a dry heat sealed package. The package is sterilized with ethylene oxide. The content is sterile unless the package is opened or damaged.

DISCLAIMER OF LIABILITY

- The manufacture will not be liable for any injury suffered to patient as a result of use of product other then intended purpose


REPORTING


Adverse reactions and potentially sight threatening complications that may reasonably be regarded as device related, need to be reported to the manufacture or local distributor

RETURN OF DAMAGED i-Cushion


Return the i-Cushion in its original package identified with the lot number and reason for returning. Do not attempt to re-sterilize the i-Cushion.


 See instruction for use


 Do not resterilize

 Do not reuse

 Do not store below 0°C & above 45°C

 Store away from sunlight

 Do not use if package damaged

 Store in dry place

STERILE EO**CE**
2460

 MFG/MD/2020/000163
i-Medical® Ophthalmic International Heidelberg GmbH
Markircher Straße 7 · D-68229 Mannheim · Germany
Fax: +49(0)621-484490-20 · www.imedical.de